VERIFICATION OF CALIBRATION REPORT
of DataMaster dmt Breath Test Instrument
State of Alaska

Serial #: 100416   **Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program**   Date:07/18/2011

External Standard Test Values

```
EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.46: 0.078
LOT #: 02511080A3
EXPIRATION: 02/01/2013

BLANK TEST              0.000   12:02
INTERNAL STANDARD       VERIFIED 12:02
EXTERNAL STANDARD       0.076   12:02
BLANK TEST              0.000   12:03
EXTERNAL STANDARD       0.076   12:03
BLANK TEST              0.000   12:04
EXTERNAL STANDARD       0.076   12:04
BLANK TEST              0.000   12:05
EXTERNAL STANDARD       0.075   12:05
BLANK TEST              0.000   12:06
EXTERNAL STANDARD       0.075   12:07
BLANK TEST              0.000   12:07

Average = 0.0756
Std Dev = 0.0005
```

Diagnostic Check

```
VERSIONS
DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C          PASSED
Breath Tube    = 48.1°C          PASSED

PUMP INFO
Flow Rate = 5.011 L/M            PASSED

DETECTOR INFO
PUMP ON                          PASSED
PUMP OFF                         PASSED

FILTER INFO
Filter 1                         PASSED
Filter 2                         PASSED
Filter 3                         PASSED

INTERNAL STANDARD                PASSED
```

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz   8/1/11
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this __1__ day of _Aug_, 20 _11_

Nikki Roth   _Nikki Roth_
Notary Public's Signature   Notary Name
My Commission Expires on _With Office_